

Limulus Amebocyte Lysate (LAL) – Report of Analysis

Test Performed	Endosafe - PTS LAL Assay
Test Date	11/19/2025
Analyst	MARIN
Instrument	Endosafe® nexgen-PTS™ (SN: 25201216)

Sample Information

Company Name	Biolongevity Labs
Sample Tested	BPC-157
Sample Lot Number	3-3-45937
Sample Concentration	10 mg
Vials/Units Received	1

Reagent Information

Cartridge Type	Endotoxin- LAL (Charles River)
Cartridge Lot #	2520301152
Cartridge Calibration Code	513635262125
Detection Range	0.05 EU/mL – 5.0 EU/mL
Spike Value	0.492 EU/mL
Spike Recovery	89% (Pass)
Test Suitability	Pass

Endotoxin Quantitation Results

Sample ID	Dilution	Replicates	Measured EU/mL	Specification	Pass/Fail
BPC-157	1x	1, 2	< 0.050 EU/mL	≤ 0.25 EU/mL	Pass

Interpretation:

Endotoxin level in **BPC-157** was measured using chromogenic LAL assay with detection at 405 nm. The result of **< 0.05 EU/mL** indicates that the endotoxin content is **below** the acceptable threshold of 0.25 EU/mL. Sample **Passes** the endotoxin limit test.

Analyst Signature:

Date: 11/19/2025

To verify the authenticity of this report, please contact **BioRegen** at
contact@bioregen.com or reference the **Report ID** and **Validation Key** provided above.

Alternatively, scan the QR code below to access the official record on our website.

